

Patient MRI Checklist

When your doctor requests an MRI scan:

- ✓ Tell your physician who requested the MRI scan that you have an implanted spinal cord stimulator device.
- ✓ Find your patient ID card. If lost, please call your Nevro HFX Care Team to request a replacement.

When making your MRI appointment:

- ✓ Provide the information below to the MRI facility when making your appointment:
 - You have a Nevro HFX spinal cord stimulator
 - If you have any other medical device implants
 - A detailed description of your device including the product number, leads, and other implant components (this information is found on your patient ID card)
 - The contact information of the physician who implanted your device
 - The Nevro HFX MRI Guidelines (available at [Nevro.com/manuals](https://www.nevro.com/manuals))

Preparing for your MRI appointment:

- ✓ Contact your Nevro HFX Care Team to confirm you can safely have the MRI scan ordered.
- ✓ Before your MRI scan, your Nevro HFX Care Team will help you prepare your device by conducting an impedance check.
 - Please refer to the Remote Control Quick Start Guide or [NevroHFX.com/resources](https://www.NevroHFX.com/resources) for step-by-step instructions.



Make sure your device is fully charged before your MRI scan.

Patient MRI Checklist

At your MRI appointment:

- ✓ Bring your patient ID card and remote control to your MRI appointment.
- ✓ Your MRI Tech will confirm your impedance check results before your MRI scan.
- ✓ Turn your stimulation OFF using your remote control.

After your MRI appointment:

- ✓ After your MRI scan, turn your stimulation back ON by pressing the Stimulation Start/Increase Button after turning your remote control ON.
- ✓ Contact your Nevro HFX Care Team to make sure your device is programmed to your pre-MRI settings.



Do you need help?

We have a dedicated team who will help you find a center that is compatible with spinal cord stimulators, help gather the information you need, and who can answer your questions.

Please call 1-844-331-1001



MRI Guidelines for Your MRI Tech

Scan the code for detailed MRI compatibility information or visit [Nevro.com/manuals](https://www.nevro.com/manuals).

Note: HFX and HFX iQ are approved for an MRI scan under certain conditions. Different models of implanted devices, including lead types, result in different approved scanning conditions. Please review the Nevro HFX MRI Guidelines carefully with your Nevro HFX Care Team before getting an MRI scan.

SENZA, SENZA II, SENZA OMNIA, OMNIA, HF10, the HF10 logo, 10 kHz Therapy, HFX, the HFX logo, HFX iQ, the HFX iQ logo, HFX Algorithm, HFX CONNECT, the HFX Connect logo, HFX ACCESS, the HFX Access logo, HFX COACH, the HFX Coach logo, HFX CLOUD, the HFX Cloud logo, RELIEF MULTIPLIED, the X logo, NEVRO, and the NEVRO logo are trademarks or registered trademarks of Nevro Corp. ©2023 Nevro Corp. All Rights Reserved.